Ĵå Dinv

DET NORSKE VERITAS CERTIFICATION AS

Compliance Testing & Field Measurement

Veritasveien 1 N-1322 HØVIK Norway Tel: +47 67 57 99 00 Fax: +47 67 57 99 11 http://www.dnv.com

RAE Systems Europe Kirstinehøj 23A

2770 Kastrup Danmark Att: Ole Olsen

Your ref.: Ole Olsen Our ref.: kra Date: 2011-09-05

STATEMENT OF TESTING

IP65 testing of MultiRAE Pro Gas Detector, version Pump, type PGM6248 according to IEC 60529 ED 2.1 (2001).

Hereby confirm that the following tests were performed at the Environmental testing laboratory of Det Norske Veritas Certification AS on the time period from 2011-07-28 to 2011-08-02:

- 1) Chapter 13.5 Special conditions for first characteristic numeral 6.
- 2) Chapter 14.2.5 Test for second characteristic numeral 5.

The detector passed the tests, see DNV test report 2011-3389.

This statement also covers all other pump versions of the gas detector; MultiRAE Pro PGM62xx with same mechanical construction.

Yours faithfully for DET NORSKE VERITAS CERTIFICATION AS

Einar Richter Jordfald Head of section Compliance Testing & Field Measurement

Knut Ranestad Ass. Consultant